Adapting to OLYMPUS 160 Series Immersible Endoscopes**



If ordering adapters:

NOTE: the ends of the adapters shown plug into the manifold.

If your machine is equipped with this type of end add an "M" when reordering adapter. e.g.: 16002M or 16006M

#16008
Specialty Channel
with Olympus Adapter (MH974)
(CLEAR TUBING)
Elevator should be moved to the half up and half down position

#16008 HP†
Specialty Channel
with Olympus Adapter
(MH974)
(CLEAR TUBING)

OR

#16026
Valve Cover

#16006
Biopsy Channel(s)
with Olympus Biopsy Adapter (MH856)
(WHITE TUBING)
Use two #16006 Biopsy Channel Adapters for dual biopsy ports.

#16008
Specialty Channel
with Olympus Adapter (MH974)
(CLEAR TUBING)

#16008 HP†
Specialty Channel
with Olympus Adapter (MH974)
(CLEAR TUBING)

OR

#16036
Water Channel Adapter
(BLUE TUBING)

#16002
Suction Channel Adapter
(GREEN TUBING)

Place waterproof cap on and leak test scope prior to adapting to system 83 manifold.

#16028
Air Channel Adapter
(YELLOW TUBING)





Biopsy Valve

Part #16030 is a filter insert that must be replaced every 6 to 8 weeks.

CAUTION: Remove all valves (air/water, suction) and biopsy caps, double biopsy, etc. consult scope manual for proper removal and cleaning procedures of components. Install certain reprocessing parts supplied by Olympus and all adapter supplied at time of In-Service. To confirm proper hook up refer to In-Service Program Report Form and/or Scope Adaptation For All Model Endoscopes Chart. If you have any questions on adapting to your scopes or if you acquire any new scopes not listed on your In-Service Program/Report Form please call Custom Ultrasonics.

** This adapter chart for Olympus Flexible Endoscopes illustrates various models and does not represent an actual model.
† For machines that are equipped with High Pressure Ports use #16008 HP Specialty Channel.

Custom Ultrasonics INCORPORATED
P.O. Box 850, Buckingham, PA 18912 (215) 364-1477 FAX (215) 364-7674 www.customultrasonics.com

FEB2003-OLM160(1)